

TSKgel®
IEC Type

TSKgel® DEAE-5PW (20)
TSKgel® DEAE-5PW (30)
TSKgel® SuperQ-5PW (20)
TSKgel® SuperQ-5PW (30)
TSKgel® SP-5PW (20)
TSKgel® SP-5PW (30)
TSKgel® SP-3PW (30)

INSTRUCTION MANUAL





TOSOH CORPORATION

Safety Precautions

To help protect you and/or your property from potential damage and ensure personal safety, please read this manual thoroughly before using the product.

[Notational Conventions]

| Notation | Explanation |
|--|--|
|  WARNING | Indicates a hazard with a medium level of risk which, if not avoided, could result in death or serious injury. |
|  CAUTION | Indicates a hazard with a low level of risk which, if not avoided, could result in minor or moderate injury. |

WARNING

■ **Keep away from fire**

Not taking proper precautions when using flammable solvents could result in fire, explosion, or poisoning.

CAUTION

■ **Use only in well-ventilated areas**

In case of insufficient ventilation, flammable and toxic solvents can cause fire, explosion, or poisoning.

■ **Do not spill solvents**

Spillage and leakage can cause fire, electric shock, poisoning, injury, or corrosion.
Wear appropriate protective gear when cleaning up a spill.

■ **Wear protective eye gear and gloves**

Organic solvents and acids should not come into direct contact with the skin.

■ **Handle the package with care**

Inappropriate handling may cause rupturing and/or splattering of the product.

■ **Only use this product for its intended use**

This product is intended for the separation and purification of small molecules and proteins. Do not use it for any other purpose.

■ **Make sure compounds are safe**

Check that the target compounds and solutions after separation and purification are safe.

■ **Proper disposal**

Dispose in accordance with local laws and regulations.

NOTE

Keep this manual with the product for future reference.

Precautions: Shipping Solvents

TSKgel® IEC Type products are shipped in an aqueous solution containing 20 % ethanol.

| | | |
|----------------------|---------------------------------------|--|
| First Aid | Inhalation | <ul style="list-style-type: none">• Move the person to an area with fresh air and rinse the mouth with plenty of water.• Call immediately for medical attention. |
| | Skin exposure | <ul style="list-style-type: none">• Wash the exposed area with plenty of soap and water. |
| | Eye exposure | <ul style="list-style-type: none">• Open the eyes as wide as possible and rinse with clean water for at least 15 minutes.• Call immediately for medical attention. |
| | Ingestion | <ul style="list-style-type: none">• Rinse the mouth with plenty of water.• Call immediately for medical attention. |
| Handling and Storage | Ventilation | <ul style="list-style-type: none">• Provide adequate air ventilation to keep organic vapor concentrations below approved level. |
| | Container handling | <ul style="list-style-type: none">• Container may break if not handled with care. |
| | Wear appropriate protective equipment | <ul style="list-style-type: none">• Use solvent-resistant gloves and protective eye gear when using this product. Use of a gas mask, additional protective clothing or rubber boots could be appropriate when handling this product. |
| | Hazardous substance storage | <ul style="list-style-type: none">• If any flammable solvents are used for shipping or storage of this product, keep away from fire or open heat sources. |
| | Storage temperature | <ul style="list-style-type: none">• Avoid storing this product at very low temperatures ($< 0\text{ }^{\circ}\text{C}$) to prevent product from freezing. |
| Waste Disposal | Disposal methods | <ul style="list-style-type: none">• Dispose in accordance with local laws and regulations. |
| | General considerations | <ul style="list-style-type: none">• Please pay attention to all safety precautions with respect to the handling and storage of this product. |

Precautions: TSKgel® Brand Chromatographic Media

| | | |
|----------------------|---------------------------------------|--|
| First Aid | Inhalation | <ul style="list-style-type: none"> • Move the person to an area with fresh air and rinse the mouth with plenty of water. • Call immediately for medical attention. |
| | Skin exposure | <ul style="list-style-type: none"> • Wash the exposed area with plenty of soap and water. |
| | Eye exposure | <ul style="list-style-type: none"> • Open the eyes as wide as possible and rinse with clean water for at least 15 minutes. • Call immediately for medical attention. |
| | Ingestion | <ul style="list-style-type: none"> • Rinse the mouth with plenty of water. • Call immediately for medical attention. |
| Handling and Storage | Ventilation | <ul style="list-style-type: none"> • Provide adequate air ventilation to keep organic vapor concentrations below approved level. |
| | Container handling | <ul style="list-style-type: none"> • Container may break if not handled with care. |
| | Wear appropriate protective equipment | <ul style="list-style-type: none"> • Use solvent-resistant gloves and protective eye gear when using this product. Use of a gas mask, additional protective clothing or rubber boots could be appropriate when handling this product. |
| | Hazardous substance storage | <ul style="list-style-type: none"> • If any flammable solvents are used for shipping or storage of this product, keep away from fire or open heat sources. |
| | Fire precautions | <ul style="list-style-type: none"> • Do not expose this chromatographic resin to fire or open heat sources. |
| Waste Disposal | Disposal methods | <ul style="list-style-type: none"> • Dispose in accordance with local laws and regulations. See below for additional precautions. |
| | General considerations | <ul style="list-style-type: none"> • Please pay attention to all safety precautions with respect to the handling and storage of this product. |
| | Disposal precaution | <ul style="list-style-type: none"> • This product can be safely incinerated. • Appropriate sulfur oxides exhaust emission precautions should be taken specifically for TSKgel® SP-5PW (20), TSKgel® SP-5PW (30) and TSKgel® SP-3PW (30). • Appropriate nitrogen oxides exhaust emission precautions should be taken specifically for TSKgel® DEAE-5PW (20), TSKgel® DEAE-5PW (30), TSKgel® SuperQ-5PW (20) and TSKgel® SuperQ-5PW (30). |

□ TSKgel® products contain combustible chromatographic packings based on a methacrylate polymer.

Table of Contents

| | |
|---------------------------------------|---|
| 1. Introduction | 1 |
| 2. Procedure for Chromatography | 1 |
| 3. Storage | 5 |
| 4. Remarks | 5 |

1. Introduction

TSKgel® IEC Type is an IEC chromatographic resin consisting of a porous and spherical polymer. TSKgel® IEC Type has the following features.

- The quantity of gel listed on the container represents the volume of gravity settled resin and not the total liquid volume.
- The change of gel volume when packed into a chromatographic column is negligible in buffers at various pH or salt concentrations.
- Applicable to fast flow-rate on column chromatography.
- Resistant for microbial growth.
- Applicable to most HPLC systems.

<Products Line-up>

| Grade | Ion Type |
|---|---------------|
| TSKgel® DEAE-5PW (20) TSKgel® DEAE-5PW (30) | Weak anion |
| TSKgel® SuperQ-5PW (20) TSKgel® SuperQ-5PW (30) | Strong anion |
| TSKgel® SP-5PW (20) TSKgel® SP-5PW (30) TSKgel® SP-3PW (30) | Strong cation |

*Note (Particle Size)
(20) : 15 μ m - 25 μ m
(30) : 20 μ m - 40 μ m

2. Procedure for Chromatography


2-1 Removal of Fines

- (1) As an example, pour the gel of 500 mL in the beaker of 3000 mL. (The capacity has six times of the gel.)
- (2) Add distilled water to a total of 2000 mL (four times of the gel) in the beaker, stir and leave until the gel settles.

| Grade | Settling Time (recommended) |
|---|-----------------------------|
| TSKgel® DEAE-5PW (30) TSKgel® SuperQ-5PW (30) TSKgel® SP-5PW (30) | 90 - 120 minutes |
| TSKgel® SP-3PW (30) | 60 - 90 minutes |
| TSKgel® DEAE-5PW (20) TSKgel® SuperQ-5PW (20) TSKgel® SP-5PW (20) | ≥120 minutes |

(3) Decant and discard the supernatant (containing fines).

(4) Repeat this process (2) and (3) at least three times.



Removal of Fines

2-2 Cleaning

TSKgel® IEC Type is shipped or stored in an aqueous solution containing 20 % ethanol.

The washing of the gel is necessary prior to use.

Pour the gel slurry on a glass filter and wash with distilled water of three times of the gel volume.

2-3 Preparation of Gel Slurry and Packing

After removing fines from the gel by decantation, wash the gel with packing solvent. The packing buffer should contain the highest salt concentration that the column will be exposed during normal use, cleaning and storage. Transfer the gel into a beaker and add the packing buffer to make an approximately 30%-50% (V/V) (recommended) slurry.

Packing the column under pressure (max. 2 MPa (recommended)) is recommended. In this case a pump and a reservoir are necessary to pack the column.

Usually the packing flow rate is at least two times faster than that of the operating flow rate. Initial packing using a gravity-settled bed can be applied, however, applying pressure from flow rate or dynamic axial compression results in the best packed columns. For this resin, best results are obtained when the packing pressure is as high as possible up to a limit of 2 MPa.

2-4 Equilibration and Performance Testing

After packing, the column should be equilibrated with 3 to 5 column volume of buffer. The column should then be tested for packing integrity using a standard performance test.

2-5 Sample Loading and Elution

The sample being purified is typically adsorbed onto the column using a lower conductivity buffer. The sample is usually desorbed from the column using with an increasing salt gradient.

2-6 Regeneration

The chromatographic resin can be regenerated after use by one of the following procedures.

2-6-1 Batch Method

Pour the gel into a beaker and suspend using an appropriate cleaning solvent. Stir and let

the gel settle for approximately 60-90 minutes. Discard the supernatant by decantation. Repeat this process 2 or 3 times.

⚠ Caution: The extremely severe cleaning method that is described below, will use HCl solution. Please note that some proteins will aggregate in acidic conditions.

* General cleaning method

First wash the gel with 0.5 mol/L-1.0 mol/L NaCl solution using the procedure mentioned above. Then equilibrate the gel with the loading buffer.

* Severe cleaning method

Wash the gel with 0.1 mol/L-0.5 mol/L NaOH followed by washing with 0.1 mol/L-0.5 mol/L NaCl solution. Then equilibrate the gel with the loading buffer.

* Extremely severe cleaning method

TSKgel® SP-5PW (20), TSKgel® SP-5PW (30), TSKgel® SP-3PW (30)

Wash the gel with 0.1 mol/L-0.5 mol/L NaOH, then water until the pH drops to near neutrality, then wash with 0.1 mol/L-0.5 mol/L HCl, and finally with 0.1 mol/L-0.5 mol/L NaCl, followed by regeneration with the loading buffer.

TSKgel® DEAE-5PW (20), TSKgel® DEAE-5PW (30),

TSKgel® SuperQ-5PW (20), TSKgel® SuperQ-5PW (30)

Wash the gel with 0.1 mol/L-0.5 mol/L HCl, then water until the pH drops to near neutrality, then wash with 0.1 mol/L-0.5 mol/L NaOH, and finally with 0.1 mol/L-0.5 mol/L NaCl, followed by regeneration with the loading buffer.

2-6-2 Column Cleaning Method

The chromatographic resin in a packed column can be regenerated easily by flowing the cleaning solvents through the column. The solvents for the column cleaning are the same as those used in the Batch Method.

[Advantages of Column Cleaning Method]

- | | |
|------------------------|--|
| * Simple Handling | Removing the gel from the column and repacking of the chromatographic resin into the column are not necessary. |
| * Good Reproducibility | Cleaning times are very consistent and reproducible. |
| * Quick Cleaning | By using a pump the cleaning times become shorter than that used by the Batch Method. |
| * Effective Cleaning | The gel can be regenerated very well with small amount of solvents compared with the Batch Method. |

3. Storage

The gel should be stored in an aqueous solution containing 20 % ethanol at ambient temperatures (4 °C -35 °C).

4. Remarks

4-1 Removal of Fines

As described in Section 2, remove fines before use. When the fines are not removed completely, there is a possibility that micro-particles may leach from column during chromatography. Leaching of the micro-particles, however, should stop after a short period of time.

4-2 Clogging of Filter

Increasing of pressure-drop or decreasing flow-rate is typically caused by filter (frit) clogging.

When this happens, remove the chromatographic resin from the column and clean the fitting and screens. Once the hardware is completely clean, repack the chromatographic resin into the column as described above.

4-3 Adsorption of Protein

When the protein or other small molecule is not adsorbed onto the column with the initial buffer, the sample should be dialyzed or desalted to reduce the conductivity. Alternatively, lower the pH of the binding buffer.

4-4 Packing Method

TOSOH recommends packing the resin into the column using a pressure-packing method.

Packing the column using a suction method or by just using gravity settling is not recommended, particularly for columns more than 10 cm in length.



TOSOH

TOSOH CORPORATION

BIOSCIENCE DIVISION

Shiba-Koen First Bldg.

3-8-2 Shiba, Minato-ku, Tokyo 105-8623, Japan

Phone: +81-3-5427-5180 Fax: +81-3-5427-5220

Web site: <http://www.separations.asia.tosohbioscience.com/>

HPLC database: www2.tosoh.co.jp/hlc/hlcdb.nsf/StartE?OpenForm

TOSOH BIOSCIENCE LLC

3604 Horizon Drive Suite 100,

King of Prussia, PA 19406, USA

Phone: +1-800-366-4875 Fax: +1-610-272-3028

E-mail: info.tbl@tosoh.com

Web site: <http://www.tosohbioscience.com/>

TOSOH BIOSCIENCE GmbH

Zettachring 6, 70567 Stuttgart, Germany

Phone: +49-711-132570 Fax: +49-711-1325789

E-mail: info.tbg@tosoh.com

Web site: <http://www.tosohbioscience.com/>

TOSOH BIOSCIENCE SHANGHAI CO., LTD.

Room 301, Plaza B, No.1289 Yi Shan Road,

Xu Hui District, Shanghai 200233, China

Phone: +86-21-3461-0856 Fax: +86-21-3461-0858

E-mail: info@tosoh.com.cn

Web site: <http://www.separations.asia.tosohbioscience.com/>

TOSOH ASIA PTE. LTD.

63 Market Street #10-03 Singapore 048942

Phone: +65-6226-5106 Fax: +65-6226-5215

E-mail: info.tsas@tosoh.com

Web site: <http://www.separations.asia.tosohbioscience.com/>

HLC, TSK-GEL, TSKgel, TSKgel SuperMultipore, BioAssist, Enantio, PStQuick, Enviropak, TOYOPEARL, ToyoScreen, TOYOPEARL GigaCap, TOYOPEARL MegaCap, TOYOPEARLPAK and TOYOPAK are registered trademarks of TOSOH CORPORATION in Japan and other countries.

This manual may not be reprinted or copied in whole or in part without written consent of TOSOH CORPORATION. The contents of the manual are subject to change without notice.

Printed in Japan